


Hydrazine
Filter End-of-Service Life Indicator
(ESLI)
(PN: 2517-2000)



Filter is good



Replace filter

Specifications



1. Application

The Hydrazine ESLI (PN: 2517-2000) is qualitative (yes/no) colorimetric indicator for the end-of-service life of filters. The indicator is designed to provide real-time indication of the end-of-service life of hydrazine filters.

2. Specifications

- | | |
|---------------------------|--|
| a. Weight: | 5g (0.18oz) |
| b. Dimensions: | 21.6mm (0.85in), ϕ : 25.0mm (1.0in) |
| c. Threading: | 0.75"-20 UNEF-2A male * |
| d. Operating temperature: | -20°C to 65°C (-4°F to 149°F) |
| e. Operating humidity: | 5% RH to 95%RH |
| Minimum detectable limit: | 0.15ppm•hr at 30 cm/sec face velocity |
| f. Color change: | Green to brown to black |
| g. Storage temperature: | 4°C to 25°C, (39°F to 77°F) |
| h. Shelf life: | 2 years at 4°C to 25°C, (39°F to 77°F) |
| i. Service life: | 2 years |

* UNEF = Unified National Imperial Screw Thread

Cross interferences and limitations: No other interference or limitation is known.

3. Operating Instructions

- Ensure that packaging pouch is intact.
- Open packaging pouch by tearing off the top part from one of side notches
- Remove ESLI from the packaging pouch.
- Remove activation plug (Figure 1).



- Hold the ESLI from the edges, as shown in Figure 2, and insert it into the ESLI housing (Figure 2).
- Screw the ESLI into the ESLI housing as shown in Figure 3.
- Replace or refill filter when the ESLI changes color brown or black.

